

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027808.D
 Acq On : 30 Mar 2022 14:17
 Operator : JC/MD
 Sample : N2041-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20220318

Quant Time: Mar 31 05:48:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	142355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61711	56.876	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.760%	
35) Dibromofluoromethane	5.391	113	47937	52.170	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.340%	
50) Toluene-d8	8.653	98	170541	52.750	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.500%	
62) 4-Bromofluorobenzene	11.079	95	64814	47.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.360%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1432	1.819	ug/l #	90
12) 1,1-Dichloroethene	2.312	96	241	0.334	ug/l #	74
30) Chloroform	5.092	83	572	0.344	ug/l #	78
44) Trichloroethene	7.129	130	322451	338.310	ug/l	90
64) Tetrachloroethene	9.275	164	476	0.524	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027808.D
Acq On : 30 Mar 2022 14:17
Operator : JC/MD
Sample : N2041-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20220318

Quant Time: Mar 31 05:48:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

