

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027811.D
 Acq On : 30 Mar 2022 15:27
 Operator : JC/MD
 Sample : N2041-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20220318

Quant Time: Mar 31 05:49:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82315	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	139182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58282	54.875	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.740%			
35) Dibromofluoromethane	5.391	113	47520	52.895	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.780%			
50) Toluene-d8	8.653	98	161724	51.164	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.320%			
62) 4-Bromofluorobenzene	11.085	95	62348	46.415	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 92.820%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	793	1.063	ug/l	81
9) 1,1,2-Trichlorotrifluo...	2.331	101	17417	22.597	ug/l	96
12) 1,1-Dichloroethene	2.325	96	2621	3.706	ug/l	94
16) Acetone	2.380	43	1007	2.142	ug/l #	78
24) 1,1-Dichloroethane	3.623	63	1392	0.951	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	3380	3.652	ug/l	96
30) Chloroform	5.099	83	2691	1.654	ug/l	94
32) 1,1,1-Trichloroethane	5.379	97	434	0.308	ug/l #	42
38) Carbon Tetrachloride	5.678	117	2107	1.685	ug/l	99
44) Trichloroethene	7.129	130	330697	354.871	ug/l	91
55) 1,1,2-Trichloroethane	9.159	97	779	0.813	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027811.D
Acq On : 30 Mar 2022 15:27
Operator : JC/MD
Sample : N2041-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20220318

Quant Time: Mar 31 05:49:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

