

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027812.D
 Acq On : 30 Mar 2022 15:51
 Operator : JC/MD
 Sample : N2041-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20220318

Quant Time: Mar 31 05:49:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	85218	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	136881	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57026	51.863	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.720%	
35) Dibromofluoromethane	5.391	113	47498	53.759	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.520%	
50) Toluene-d8	8.653	98	162775	52.362	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.720%	
62) 4-Bromofluorobenzene	11.079	95	64118	48.535	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.060%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	427	0.535	ug/l	96
44) Trichloroethene	7.129	130	29572	32.267	ug/l	85
64) Tetrachloroethene	9.275	164	1312	1.569	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027812.D
Acq On : 30 Mar 2022 15:51
Operator : JC/MD
Sample : N2041-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20220318

Quant Time: Mar 31 05:49:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

