

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027814.D
 Acq On : 30 Mar 2022 16:38
 Operator : JC/MD
 Sample : N2041-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20220317DL

Quant Time: Mar 31 05:50:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72212	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	127824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62033	66.578	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	133.160%
35) Dibromofluoromethane	5.391	113	43266	52.439	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.880%
50) Toluene-d8	8.653	98	158484	54.594	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.180%
62) 4-Bromofluorobenzene	11.079	95	63421	51.409	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.820%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	2308	3.413	ug/l	98
12) 1,1-Dichloroethene	2.319	96	620	0.999	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	805	0.991	ug/l	82
44) Trichloroethene	7.129	130	52560	61.414	ug/l	92
64) Tetrachloroethene	9.275	164	1099	1.335	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027814.D
Acq On : 30 Mar 2022 16:38
Operator : JC/MD
Sample : N2041-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20220317DL

Quant Time: Mar 31 05:50:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

