

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027815.D
 Acq On : 30 Mar 2022 17:01
 Operator : JC/MD
 Sample : N2074-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-20220318

Quant Time: Mar 31 05:50:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	144225	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119290	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54190	50.217	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.440%
35) Dibromofluoromethane	5.391	113	47591	51.122	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.240%
50) Toluene-d8	8.653	98	161227	49.223	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.086	95	61921	44.485	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.980%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	746	1.562	ug/l	88
44) Trichloroethene	7.123	130	672	0.696	ug/l	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027815.D
Acq On : 30 Mar 2022 17:01
Operator : JC/MD
Sample : N2074-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20220318

Quant Time: Mar 31 05:50:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

