

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027816.D
 Acq On : 30 Mar 2022 17:24
 Operator : JC/MD
 Sample : N2074-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D1-20220318

Quant Time: Mar 31 05:50:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	82236	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	133502	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55358	52.172	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.340%	
35) Dibromofluoromethane	5.391	113	46896	54.421	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.840%	
50) Toluene-d8	8.652	98	159845	52.721	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.440%	
62) 4-Bromofluorobenzene	11.079	95	62194	48.270	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.540%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	566	0.760	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.330	101	1129	1.466	ug/l	99
12) 1,1-Dichloroethene	2.324	96	647	0.916	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	602	0.651	ug/l	87
30) Chloroform	5.098	83	817	0.503	ug/l	99
44) Trichloroethene	7.128	130	81275	90.927	ug/l	85
55) 1,1,2-Trichloroethane	9.158	97	325	0.354	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027816.D
Acq On : 30 Mar 2022 17:24
Operator : JC/MD
Sample : N2074-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D1-20220318

Quant Time: Mar 31 05:50:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

