

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027820.D
 Acq On : 30 Mar 2022 18:58
 Operator : JC/MD
 Sample : N2041-15DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20220317DL

Quant Time: Mar 31 05:51:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	79044	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	134014	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118650	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57694	56.569	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.140%
35) Dibromofluoromethane	5.391	113	47006	54.340	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.680%
50) Toluene-d8	8.653	98	156989	51.581	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.160%
62) 4-Bromofluorobenzene	11.085	95	61956	47.902	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.800%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	372	0.503	ug/l	# 91
44) Trichloroethene	7.129	130	46769	52.123	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027820.D
Acq On : 30 Mar 2022 18:58
Operator : JC/MD
Sample : N2041-15DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20220317DL

Quant Time: Mar 31 05:51:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

