

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027827.D
 Acq On : 30 Mar 2022 22:30
 Operator : JC/MD
 Sample : VX0330WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0330WBL02

Quant Time: Mar 31 07:30:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75690	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	134843	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52383	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64415	65.958	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	131.920%
35) Dibromofluoromethane	5.391	113	46589	53.527	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.060%
50) Toluene-d8	8.653	98	164265	53.640	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.280%
62) 4-Bromofluorobenzene	11.085	95	64195	49.328	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027827.D
Acq On : 30 Mar 2022 22:30
Operator : JC/MD
Sample : VX0330WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0330WBL02

Quant Time: Mar 31 07:30:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

