

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027831.D
 Acq On : 31 Mar 2022 00:04
 Operator : JC/MD
 Sample : N2041-18DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20220317DL

Quant Time: Mar 31 07:32:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	130905	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61873	61.090	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 122.180%	
35) Dibromofluoromethane	5.391	113	46314	54.812	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.620%	
50) Toluene-d8	8.653	98	155591	52.336	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.680%	
62) 4-Bromofluorobenzene	11.079	95	60759	48.092	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.180%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	220	0.309	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.331	101	2068	2.814	ug/l	95
12) 1,1-Dichloroethene	2.325	96	730	1.082	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	367	0.416	ug/l	85
30) Chloroform	5.105	83	530	0.342	ug/l	98
38) Carbon Tetrachloride	5.678	117	478	0.407	ug/l	98
44) Trichloroethene	7.129	130	40976	46.752	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027831.D
Acq On : 31 Mar 2022 00:04
Operator : JC/MD
Sample : N2041-18DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20220317DL

Quant Time: Mar 31 07:32:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

