

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX033022\
 Data File : VX027826.D
 Acq On : 30 Mar 2022 21:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 07:29:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	51.230	-2.5	87	0.00
3 P	Chloromethane	50.000	46.704	6.6	88	0.00
4 C	Vinyl Chloride	50.000	49.157	1.7#	90	0.00
5 T	Bromomethane	50.000	44.951	10.1	91	0.00
6 T	Chloroethane	50.000	47.424	5.2	94	0.00
7 T	Trichlorofluoromethane	50.000	47.942	4.1	89	0.00
8 T	Diethyl Ether	50.000	49.905	0.2	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.956	2.1	93	0.00
10 T	Methyl Iodide	50.000	48.091	3.8	88	0.00
11 T	Tert butyl alcohol	250.000	279.482	-11.8	102	0.00
12 CM	1,1-Dichloroethene	50.000	46.649	6.7#	93	0.00
13 T	Acrolein	250.000	296.916	-18.8	110	0.00
14 T	Allyl chloride	50.000	50.089	-0.2	96	0.00
15 T	Acrylonitrile	250.000	271.471	-8.6	98	0.00
16 T	Acetone	250.000	258.291	-3.3	98	0.00
17 T	Carbon Disulfide	50.000	40.493	19.0	83	0.00
18 T	Methyl Acetate	50.000	53.383	-6.8	102	0.00
19 T	Methyl tert-butyl Ether	50.000	52.673	-5.3	97	0.00
20 T	Methylene Chloride	50.000	46.449	7.1	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.106	11.8	87	0.00
22 T	Diisopropyl ether	50.000	52.299	-4.6	108	0.00
23 T	Vinyl Acetate	250.000	270.259	-8.1	110	0.00
24 P	1,1-Dichloroethane	50.000	50.596	-1.2	104	0.00
25 T	2-Butanone	250.000	277.741	-11.1	115	0.00
26 T	2,2-Dichloropropane	50.000	46.868	6.3	94	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.402	1.2	94	0.00
28 T	Bromochloromethane	50.000	50.402	-0.8	104	0.00
29 T	Tetrahydrofuran	250.000	273.099	-9.2	112	0.00
30 C	Chloroform	50.000	50.750	-1.5#	101	0.00
31 T	Cyclohexane	50.000	47.622	4.8	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.167	-2.3	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	43.927	12.1	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	43.712	12.6	92	0.00
36 T	1,1-Dichloropropene	50.000	47.750	4.5	96	0.00
37 T	Ethyl Acetate	50.000	51.435	-2.9	114	0.00
38 T	Carbon Tetrachloride	50.000	49.898	0.2	98	0.00
39 T	Methylcyclohexane	50.000	43.827	12.3	88	0.00
40 TM	Benzene	50.000	47.142	5.7	92	0.00
41 T	Methacrylonitrile	50.000	52.720	-5.4	109	0.00
42 TM	1,2-Dichloroethane	50.000	47.769	4.5	88	0.00
43 T	Isopropyl Acetate	50.000	49.604	0.8	88	0.00
44 TM	Trichloroethene	50.000	46.254	7.5	94	0.00
45 C	1,2-Dichloropropane	50.000	47.624	4.8#	92	0.00
46 T	Dibromomethane	50.000	46.709	6.6	91	0.00
47 T	Bromodichloromethane	50.000	42.925	14.2	83	0.00
48 T	Methyl methacrylate	50.000	47.318	5.4	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027826.D
 Acq On : 30 Mar 2022 21:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 07:29:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1034.798	-3.5	97	0.00
50 S	Toluene-d8	50.000	41.452	17.1	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	213.519	14.6	80	0.00
52 CM	Toluene	50.000	47.324	5.4#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	46.267	7.5	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	44.737	10.5	85	0.00
55 T	1,1,2-Trichloroethane	50.000	46.400	7.2	92	0.00
56 T	Ethyl methacrylate	50.000	48.249	3.5	91	0.00
57 T	1,3-Dichloropropane	50.000	46.932	6.1	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	221.452	11.4	82	0.00
59 T	2-Hexanone	250.000	251.122	-0.4	94	0.00
60 T	Dibromochloromethane	50.000	47.607	4.8	91	0.00
61 T	1,2-Dibromoethane	50.000	46.442	7.1	90	0.00
62 S	4-Bromofluorobenzene	50.000	40.794	18.4	83	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	47.872	4.3	91	0.00
65 PM	Chlorobenzene	50.000	46.702	6.6	89	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.013	2.0	90	0.00
67 C	Ethyl Benzene	50.000	47.841	4.3#	88	0.00
68 T	m/p-Xylenes	100.000	95.962	4.0	89	0.00
69 T	o-Xylene	50.000	48.583	2.8	88	0.00
70 T	Styrene	50.000	48.826	2.3	89	0.00
71 P	Bromoform	50.000	46.351	7.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	46.571	6.9	89	0.00
74 T	N-ethyl acetate	50.000	49.868	0.3	90	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.771	2.5	92	0.00
76 T	1,2,3-Trichloropropane	50.000	50.791	-1.6	92	0.00
77 T	Bromobenzene	50.000	46.749	6.5	88	0.00
78 T	n-propylbenzene	50.000	46.257	7.5	89	0.00
79 T	2-Chlorotoluene	50.000	45.740	8.5	89	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.789	6.4	89	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.648	10.7	86	0.00
82 T	4-Chlorotoluene	50.000	46.427	7.1	89	0.00
83 T	tert-Butylbenzene	50.000	46.425	7.2	90	0.00
84 T	1,2,4-Trimethylbenzene	50.000	46.179	7.6	88	0.00
85 T	sec-Butylbenzene	50.000	47.343	5.3	89	0.00
86 T	p-Isopropyltoluene	50.000	46.957	6.1	87	0.00
87 T	1,3-Dichlorobenzene	50.000	45.151	9.7	88	0.00
88 T	1,4-Dichlorobenzene	50.000	44.148	11.7	88	0.00
89 T	n-Butylbenzene	50.000	46.513	7.0	86	0.00
90 T	Hexachloroethane	50.000	46.317	7.4	88	0.00
91 T	1,2-Dichlorobenzene	50.000	46.375	7.2	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.228	-2.5	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	44.679	10.6	88	0.00
94 T	Hexachlorobutadiene	50.000	45.293	9.4	87	0.00
95 T	Naphthalene	50.000	47.338	5.3	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
Data File : VX027826.D
Acq On : 30 Mar 2022 21:42
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 31 07:29:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	46.140	7.7	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6