


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027797.D
 Acq On : 30 Mar 2022 09:40
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Title : SW846 8260
 Last Update : Wed Mar 30 04:39:27 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	3037	PASS
75	95	30	60	50.9	9920	PASS
95	95	100	100	100.0	19491	PASS
96	95	5	9	6.3	1228	PASS
173	174	0.00	2	0.7	96	PASS
174	95	50	100	71.5	13939	PASS
175	174	5	9	7.3	1020	PASS
176	174	95	101	96.8	13490	PASS
177	176	5	9	6.3	849	PASS