



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033022\
 Data File : VX027825.D
 Acq On : 30 Mar 2022 20:55
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Title : SW846 8260
 Last Update : Wed Mar 30 04:39:27 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	2882	PASS
75	95	30	60	51.6	8601	PASS
95	95	100	100	100.0	16668	PASS
96	95	5	9	6.9	1157	PASS
173	174	0.00	2	0.8	100	PASS
174	95	50	100	75.8	12636	PASS
175	174	5	9	6.8	864	PASS
176	174	95	101	97.2	12281	PASS
177	176	5	9	6.7	819	PASS