

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
 Data File : VX034851.D
 Acq On : 30 Mar 2023 14:42
 Operator : JC/MD
 Sample : 02091-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Mar 31 01:32:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 11:02:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	415412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	365937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	176125	47.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.420%
35) Dibromofluoromethane	5.385	113	131388	47.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.640%
50) Toluene-d8	8.647	98	503839	49.680	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	179413	45.716	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	81950	54.852	ug/l	97
18) Methyl Acetate	2.703	43	1470	0.279	ug/l	91
25) 2-Butanone	4.568	43	25372	11.702	ug/l	92
51) 4-Methyl-2-Pentanone	8.574	43	38124	9.102	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
Data File : VX034851.D
Acq On : 30 Mar 2023 14:42
Operator : JC/MD
Sample : 02091-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Mar 31 01:32:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 11:02:25 2023
Response via : Initial Calibration

