

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
 Data File : VX034852.D
 Acq On : 30 Mar 2023 15:05
 Operator : JC/MD
 Sample : 02074-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-032323

Quant Time: Mar 31 01:32:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 11:02:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	252716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	431940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	384763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	179158	46.355	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.700%
35) Dibromofluoromethane	5.385	113	135218	47.330	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.647	98	524999	49.786	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.580%
62) 4-Bromofluorobenzene	11.079	95	188253	46.133	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.260%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	2382	0.766	ug/l #	88
16) Acetone	2.380	43	4127	2.639	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	8585	2.599	ug/l	93
30) Chloroform	5.099	83	10526	1.905	ug/l	95
44) Trichloroethene	7.129	130	2778	0.897	ug/l	93
64) Tetrachloroethene	9.275	164	8838	3.217	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
Data File : VX034852.D
Acq On : 30 Mar 2023 15:05
Operator : JC/MD
Sample : 02074-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-032323

Quant Time: Mar 31 01:32:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 11:02:25 2023
Response via : Initial Calibration

