

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
 Data File : VX034855.D
 Acq On : 30 Mar 2023 16:14
 Operator : JC/MD
 Sample : 02074-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWCB-05-40-45-032323-FD

Quant Time: Mar 31 01:33:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 11:02:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	418277	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	375235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159882	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	175977	47.624	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.240%
35) Dibromofluoromethane	5.385	113	131932	47.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.647	98	508944	49.840	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.680%
62) 4-Bromofluorobenzene	11.079	95	178652	45.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
Data File : VX034855.D
Acq On : 30 Mar 2023 16:14
Operator : JC/MD
Sample : 02074-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWCB-05-40-45-032323-FD

Quant Time: Mar 31 01:33:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 11:02:25 2023
Response via : Initial Calibration

