

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
 Data File : VX034859.D
 Acq On : 30 Mar 2023 17:47
 Operator : JC/MD
 Sample : 02096-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Mar 31 01:35:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 11:02:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	236231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	408297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160774	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169283	46.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.720%
35) Dibromofluoromethane	5.385	113	127208	47.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	488665	49.024	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.040%
62) 4-Bromofluorobenzene	11.079	95	171617	44.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
Data File : VX034859.D
Acq On : 30 Mar 2023 17:47
Operator : JC/MD
Sample : 02096-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Mar 31 01:35:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 11:02:25 2023
Response via : Initial Calibration

