

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
 Data File : VX034860.D
 Acq On : 30 Mar 2023 18:10
 Operator : JC/MD
 Sample : 02096-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Mar 31 01:35:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 30 11:02:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	230126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	399279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	354571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157500	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	168432	47.858	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.720%
35) Dibromofluoromethane	5.385	113	124051	46.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.940%
50) Toluene-d8	8.647	98	479801	49.222	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.079	95	172967	45.854	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033023\
Data File : VX034860.D
Acq On : 30 Mar 2023 18:10
Operator : JC/MD
Sample : 02096-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Mar 31 01:35:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 30 11:02:25 2023
Response via : Initial Calibration

