

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033120\
 Data File : VX015449.D
 Acq On : 31 Mar 2020 13:27
 Operator : JC/SP
 Sample : L2042-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-01-(SB-04)

Quant Time: Apr 01 08:04:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1104030	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1813622	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1733269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	937308	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	762013	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
35) Dibromofluoromethane	5.47	113	570305	49.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.38%	
50) Toluene-d8	8.70	98	2212473	51.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.12	95	953063	54.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.88%	
Target Compounds						
16) Acetone	2.42	43	19172	2.866	ug/l	98
17) Carbon Disulfide	2.56	76	42741	1.344	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033120\
Data File : VX015449.D
Acq On : 31 Mar 2020 13:27
Operator : JC/SP
Sample : L2042-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-01-(SB-04)

Quant Time: Apr 01 08:04:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

