

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033120\
 Data File : VX015454.D
 Acq On : 31 Mar 2020 15:20
 Operator : JC/SP
 Sample : PB127897ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127897ZHE#09

Quant Time: Apr 01 08:22:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1066687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1774512	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1709888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	921108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	740670	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
35) Dibromofluoromethane	5.47	113	559351	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
50) Toluene-d8	8.70	98	2121839	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.12	95	926282	54.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.16%	
Target Compounds						
16) Acetone	2.42	43	42059	6.507	ug/l	99
20) Methylene Chloride	2.84	84	16243	1.345	ug/l	94
43) Isopropyl Acetate	6.42	43	403764	11.235	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033120\
Data File : VX015454.D
Acq On : 31 Mar 2020 15:20
Operator : JC/SP
Sample : PB127897ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127897ZHE#09

Quant Time: Apr 01 08:22:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

