

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033120\
 Data File : VX015444.D
 Acq On : 31 Mar 2020 11:13
 Operator : JC/SP
 Sample : VX0331MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0331MBL01

Quant Time: Apr 01 07:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1107148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1776631	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1624205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	828515	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	734449	47.03	ug/l	0.00
Spiked Amount						
			Recovery	=	94.06%	
35) Dibromofluoromethane	5.47	113	568065	50.02	ug/l	0.00
Spiked Amount						
			Recovery	=	100.04%	
50) Toluene-d8	8.70	98	2162576	51.72	ug/l	0.00
Spiked Amount						
			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.12	95	890023	52.38	ug/l	0.00
Spiked Amount						
			Recovery	=	104.76%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033120\
Data File : VX015444.D
Acq On : 31 Mar 2020 11:13
Operator : JC/SP
Sample : VX0331MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0331MBL01

Quant Time: Apr 01 07:41:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 778

Delta R.T. 0.00 min

Lab File: VX015444.D

Acq: 31 Mar 2020 11:13

Instrument :

MSVOA_X

ClientSampleId :


VX0331MBL01

Tot Ion:168 Resp: 1107148

Ion Ratio Lower Upper




168 100

99 57.5 46.4 69.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 47.035 ug/l

RT: 6.03 min Scan# 844

Delta R.T. 0.00 min

Lab File: VX015444.D


Acq: 31 Mar 2020 11:13

Tgt Ion: 65 Resp: 734449

Ion Ratio Lower Upper



65 100


67 50.5 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 975

Delta R.T. -0.01 min

Lab File: VX015444.D

Acq: 31 Mar 2020 11:13

Instrument :

MSVOA_X

ClientSampleId :

VX0331MBL01


Tot Ion:114 Resp: 1776631

Ion Ratio Lower Upper

114 100

63 23.5 0.0 44.2




88 16.6 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.020 ug/l

RT: 5.47 min Scan# 751

Delta R.T. -0.01 min

Lab File: VX015444.D

Acq: 31 Mar 2020 11:13


Tgt Ion:113 Resp: 568065

Ion Ratio Lower Upper

113 100

111 104.6 81.5 122.3



192 21.5 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.721 ug/l

RT: 8.70 min Scan# 1281

Delta R.T. 0.00 min

Lab File: VX015444.D

Acq: 31 Mar 2020 11:13

Instrument :

MSVOA_X

ClientSampleId :





VX0331MBL01

Tot Ion: 98 Resp: 2162576

Ion Ratio Lower Upper

98 100

100 65.9 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 52.377 ug/l

RT: 11.12 min Scan# 1679

Delta R.T. 0.00 min

Lab File: VX015444.D

Acq: 31 Mar 2020 11:13




Tgt Ion: 95 Resp: 890023


Ion Ratio Lower Upper

95 100

174 77.0 0.0 158.6

176 76.2 0.0 159.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1511
Delta R.T. 0.00 min
Lab File: VX015444.D
Acq: 31 Mar 2020 11:13

Instrument :
MSVOA_X
ClientSampled :
VX0331MBL01

Tot Ion:117 Resp: 1624205
Ion Ratio Lower Upper
117 100
82 59.5 47.8 71.8
119 32.9 26.8 40.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1833
Delta R.T. 0.00 min
Lab File: VX015444.D
Acq: 31 Mar 2020 11:13

Tgt Ion:152 Resp: 828515
Ion Ratio Lower Upper
152 100
115 58.8 39.0 116.9
150 158.1 0.0 340.8

