

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX033120\
 Data File : VX015455.D
 Acq On : 31 Mar 2020 15:42
 Operator : JC/SP
 Sample : PB127897ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127897ZHE#10

Quant Time: Apr 01 08:25:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1143053	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1850605	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1788323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	941776	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	776539	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
35) Dibromofluoromethane	5.47	113	591683	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
50) Toluene-d8	8.70	98	2232261	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
62) 4-Bromofluorobenzene	11.12	95	966249	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
Target Compounds						
16) Acetone	2.42	43	41118	5.937	ug/l	96
20) Methylene Chloride	2.84	84	13883	1.073	ug/l	97
43) Isopropyl Acetate	6.42	43	355133	9.475	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX033120\
Data File : VX015455.D
Acq On : 31 Mar 2020 15:42
Operator : JC/SP
Sample : PB127897ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127897ZHE#10

Quant Time: Apr 01 08:25:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 779

Delta R.T. 0.01 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42

Instrument :

MSVOA_X

ClientSampleId :


PB127897ZHE#10

Tot Ion:168 Resp: 1143053

Ion Ratio Lower Upper

168 100

99 55.4 46.4 69.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.64

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 5.937 ug/l

RT: 2.42 min Scan# 251

Delta R.T. 0.00 min

Lab File: VX015455.D


Acq: 31 Mar 2020 15:42

Tgt Ion: 43 Resp: 41118

Ion Ratio Lower Upper

43 100

58 29.7 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.42

25000

20000


15000


10000

5000

0

Time-->









#20
Methylene Chloride
Concen: 1.073 ug/l
RT: 2.84 min Scan# 320
Delta R.T. 0.01 min
Lab File: VX015455.D
Acq: 31 Mar 2020 15:42




Instrument :
MSVOA_X
ClientSampleId :
PB127897ZHE#10


Tot Ion:	84	Resp:	13883
Ion	Ratio	Lower	Upper
84	100		
49	138.0	111.3	166.9
51	35.9	34.0	51.0
86	62.2	52.5	78.7



#33
1,2-Dichloroethane-d4
Concen: 48.168 ug/l
RT: 6.04 min Scan# 845
Delta R.T. 0.01 min
Lab File: VX015455.D
Acq: 31 Mar 2020 15:42

Tgt Ion:	65	Resp:	776539
Ion	Ratio	Lower	Upper
65	100		
67	50.7	0.0	102.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.84 min Scan# 976

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42

Instrument :

MSVOA_X

ClientSampleId :

PB127897ZHE#10


Tot Ion:114 Resp: 1850605

Ion Ratio Lower Upper

114 100

63 22.0 0.0 44.2




88 15.9 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.017 ug/l

RT: 5.47 min Scan# 752

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42


Tgt Ion:113 Resp: 591683

Ion Ratio Lower Upper

113 100

111 101.8 81.5 122.3



192 21.4 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 9.475 ug/l

RT: 6.42 min Scan# 907

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42

Instrument :

MSVOA_X

ClientSampleId :

PB127897ZHE#10


Tot Ion: 43 Resp: 355133

Ion Ratio Lower Upper

43 100

61 18.6 14.8 22.2




87 11.8 9.4 14.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#50

Toluene-d8

Concen: 51.254 ug/l

RT: 8.70 min Scan# 1281

Delta R.T. 0.00 min

Lab File: VX015455.D


Acq: 31 Mar 2020 15:42

Tgt Ion: 98 Resp: 2232261

Ion Ratio Lower Upper



98 100


100 66.4 53.5 80.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#62

4-Bromofluorobenzene

Concen: 54.590 ug/l

RT: 11.12 min Scan# 1679

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42

Instrument :

MSVOA_X

ClientSampleId :

PB127897ZHE#10


Tot Ion: 95 Resp: 966249

Ion Ratio Lower Upper

95 100

174 77.7 0.0 158.6




176 76.1 0.0 159.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1511

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42


Tgt Ion: 117 Resp: 1788323

Ion Ratio Lower Upper

117 100

82 56.2 47.8 71.8



119 30.7 26.8 40.2




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VX015455.D

Acq: 31 Mar 2020 15:42

Instrument :

MSVOA_X

ClientSampleId :

PB127897ZHE#10


Tot Ion:152 Resp: 941776

Ion Ratio Lower Upper

152 100

115 59.6 39.0 116.9

150 155.5 0.0 340.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

