

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021676.D
 Acq On : 31 Mar 2021 12:04
 Operator : JC/MD
 Sample : M1842-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20210323DL

Quant Time: Apr 01 03:22:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	102527	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	103248	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	41781	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	48857	51.35	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.70%
35) Dibromofluoromethane	5.470	113	37774	51.87	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.74%
50) Toluene-d8	8.700	98	130836	50.41	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.82%
62) 4-Bromofluorobenzene	11.118	95	49140	48.70	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.40%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.370	101	363	0.62	ug/l	97
44) Trichloroethene	7.188	130	41855	55.60	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021676.D
Acq On : 31 Mar 2021 12:04
Operator : JC/MD
Sample : M1842-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20210323DL

Quant Time: Apr 01 03:22:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

