

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021678.D
 Acq On : 31 Mar 2021 12:45
 Operator : JC/MD
 Sample : M1842-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D1-20210324

Quant Time: Apr 01 03:23:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	58356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	105714	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	50057	51.55	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.10%	
35) Dibromofluoromethane	5.470	113	39423	52.50	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.00%	
50) Toluene-d8	8.694	98	134688	50.33	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.66%	
62) 4-Bromofluorobenzene	11.118	95	52982	50.93	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.86%	
Target Compounds						
16) Acetone	2.423	43	1806	4.92	ug/l	89
44) Trichloroethene	7.188	130	6215	8.01	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021678.D
Acq On : 31 Mar 2021 12:45
Operator : JC/MD
Sample : M1842-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D1-20210324

Quant Time: Apr 01 03:23:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

