

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021681.D
 Acq On : 31 Mar 2021 13:48
 Operator : JC/MD
 Sample : M1842-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE139D2-20210324

Quant Time: Apr 10 01:06:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107418	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47946	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	49737	51.30	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.60%	
35) Dibromofluoromethane	5.470	113	39217	51.40	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.80%	
50) Toluene-d8	8.694	98	134866	49.60	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.20%	
62) 4-Bromofluorobenzene	11.118	95	54839	51.88	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.76%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	1823	3.04	ug/l	96
12) 1,1-Dichloroethene	2.358	96	1866	3.18	ug/l	94
16) Acetone	2.423	43	1993	5.44	ug/l #	84
27) cis-1,2-Dichloroethene	4.570	96	1462	1.95	ug/l	93
38) Carbon Tetrachloride	5.753	117	2219	1.96	ug/l	92
44) Trichloroethene	7.188	130	150147	190.37	ug/l	92
55) 1,1,2-Trichloroethane	9.188	97	523	0.67	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021681.D
Acq On : 31 Mar 2021 13:48
Operator : JC/MD
Sample : M1842-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE139D2-20210324

Quant Time: Apr 10 01:06:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

