

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX033121\  
 Data File : VX021683.D  
 Acq On : 31 Mar 2021 14:33  
 Operator : JC/MD  
 Sample : M1842-19  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE139D1-20210324

Quant Time: Apr 01 03:25:13 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032421W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 24 12:25:26 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units     | Dev(Min) |
|------------------------------|--------|-------|----------|----------|-----------|----------|
| -----                        |        |       |          |          |           |          |
| Internal Standards           |        |       |          |          |           |          |
| 1) Pentafluorobenzene        | 5.635  | 168   | 56016    | 50.00    | ug/l      | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.835  | 114   | 105000   | 50.00    | ug/l      | 0.00     |
| 63) Chlorobenzene-d5         | 10.094 | 117   | 112115   | 50.00    | ug/l      | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.059 | 152   | 50527    | 50.00    | ug/l      | 0.00     |
| System Monitoring Compounds  |        |       |          |          |           |          |
| 33) 1,2-Dichloroethane-d4    | 6.035  | 65    | 49691    | 53.31    | ug/l      | 0.00     |
| Spiked Amount                | 50.000 | Range | 78 - 117 | Recovery | = 106.62% |          |
| 35) Dibromofluoromethane     | 5.470  | 113   | 37941    | 50.87    | ug/l      | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | = 101.74% |          |
| 50) Toluene-d8               | 8.694  | 98    | 133577   | 50.26    | ug/l      | 0.00     |
| Spiked Amount                | 50.000 | Range | 92 - 112 | Recovery | = 100.52% |          |
| 62) 4-Bromofluorobenzene     | 11.118 | 95    | 55488    | 53.70    | ug/l      | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | = 107.40% |          |
| Target Compounds             |        |       |          |          |           |          |
|                              |        |       |          |          |           | Qvalue   |
| 9) 1,1,2-Trichlorotrifluo... | 2.370  | 101   | 3155     | 5.47     | ug/l      | 95       |
| 12) 1,1-Dichloroethene       | 2.358  | 96    | 2179     | 3.86     | ug/l      | 82       |
| 16) Acetone                  | 2.423  | 43    | 1827     | 5.18     | ug/l      | 95       |
| 27) cis-1,2-Dichloroethene   | 4.564  | 96    | 1498     | 2.08     | ug/l      | 98       |
| 30) Chloroform               | 5.182  | 83    | 1258     | 0.99     | ug/l      | 98       |
| 38) Carbon Tetrachloride     | 5.753  | 117   | 1571     | 1.42     | ug/l      | 88       |
| 44) Trichloroethene          | 7.188  | 130   | 130747   | 169.59   | ug/l      | 98       |
| 55) 1,1,2-Trichloroethane    | 9.194  | 97    | 939      | 1.23     | ug/l      | 99       |
| -----                        |        |       |          |          |           |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX033121\  
Data File : VX021683.D  
Acq On : 31 Mar 2021 14:33  
Operator : JC/MD  
Sample : M1842-19  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE139D1-20210324

Quant Time: Apr 01 03:25:13 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032421W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 24 12:25:26 2021  
Response via : Initial Calibration

