

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021684.D
 Acq On : 31 Mar 2021 14:53
 Operator : JC/MD
 Sample : M1842-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20210324

Quant Time: Apr 01 03:25:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	55749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	104605	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110913	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	49844	53.74	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.48%	
35) Dibromofluoromethane	5.464	113	37558	50.55	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.10%	
50) Toluene-d8	8.700	98	133134	50.28	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.56%	
62) 4-Bromofluorobenzene	11.124	95	53703	52.17	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.34%	
Target Compounds						
16) Acetone	2.423	43	1501	4.28	ug/l #	83
19) Methyl tert-butyl Ether	3.170	73	2940	1.40	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021684.D
Acq On : 31 Mar 2021 14:53
Operator : JC/MD
Sample : M1842-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20210324

Quant Time: Apr 01 03:25:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

