

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021685.D
 Acq On : 31 Mar 2021 15:14
 Operator : JC/MD
 Sample : M1842-18DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20210324DL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2021 4:13:43 PM

Quant Time: Apr 01 03:25:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	57202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	103302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50673	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	49456	51.96	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.92%		
35) Dibromofluoromethane	5.470	113	38678	52.71	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.42%		
50) Toluene-d8	8.694	98	133549	51.07	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.14%		
62) 4-Bromofluorobenzene	11.118	95	55765	54.85	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.70%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.358	96	243	0.42	ug/l	95
27) cis-1,2-Dichloroethene	4.564	96	3594	4.89	ug/l	95
44) Trichloroethene	7.188	130	32843	43.30	ug/l	95
45) 1,2-Dichloropropane	7.494	63	3189	4.10	ug/l #	87
64) Tetrachloroethene	9.318	164	1537	2.07	ug/l	92
76) 1,2,3-Trichloropropane	11.277	75	472m	0.40	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021685.D
Acq On : 31 Mar 2021 15:14
Operator : JC/MD
Sample : M1842-18DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20210324DL

Manual Integrations
APPROVED

MMDadoda
4/1/2021 4:13:43 PM

Quant Time: Apr 01 03:25:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

