

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021690.D
 Acq On : 31 Mar 2021 17:01
 Operator : JC/MD
 Sample : M1842-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20210324

Quant Time: Apr 01 03:27:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	55443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	98723	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	103416	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	43654	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	48302	52.36	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.72%	
35) Dibromofluoromethane	5.464	113	36674	52.30	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.60%	
50) Toluene-d8	8.694	98	124800	49.94	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.88%	
62) 4-Bromofluorobenzene	11.118	95	50165	51.63	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.26%	
Target Compounds						
16) Acetone	2.423	43	1731	4.96	ug/l	Qvalue # 84
19) Methyl tert-butyl Ether	3.164	73	3068	1.47	ug/l	93
44) Trichloroethene	7.188	130	581	0.80	ug/l	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021690.D
Acq On : 31 Mar 2021 17:01
Operator : JC/MD
Sample : M1842-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20210324

Quant Time: Apr 01 03:27:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

