

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021691.D
 Acq On : 31 Mar 2021 17:24
 Operator : JC/MD
 Sample : M1843-21DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20210324DL

Quant Time: Apr 01 03:27:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	53072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	99236	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	101536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46101	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	47167	53.41	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.82%	
35) Dibromofluoromethane	5.470	113	36374	51.60	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.20%	
50) Toluene-d8	8.694	98	126070	50.19	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.38%	
62) 4-Bromofluorobenzene	11.118	95	50610	51.82	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.64%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.364	101	254	0.46	ug/l	Qvalue 89
44) Trichloroethene	7.188	130	43789	60.10	ug/l	100
64) Tetrachloroethene	9.318	164	269	0.40	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021691.D
Acq On : 31 Mar 2021 17:24
Operator : JC/MD
Sample : M1843-21DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20210324DL

Quant Time: Apr 01 03:27:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

