

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021694.D
 Acq On : 31 Mar 2021 18:34
 Operator : JC/MD
 Sample : PB135444TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135444TB

Quant Time: Apr 01 03:29:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	58447	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49499	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	50154	51.57	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.14%			
35) Dibromofluoromethane	5.464	113	39120	50.41	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.82%			
50) Toluene-d8	8.694	98	140241	50.71	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.42%			
62) 4-Bromofluorobenzene	11.118	95	54985	51.14	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.28%			
Target Compounds					Qvalue	
16) Acetone	2.423	43	18169	49.41	ug/l #	86
20) Methylene Chloride	2.834	84	5416	7.55	ug/l #	83
25) 2-Butanone	4.640	43	3523	6.32	ug/l #	87
43) Isopropyl Acetate	6.411	43	8735	4.35	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021694.D
Acq On : 31 Mar 2021 18:34
Operator : JC/MD
Sample : PB135444TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135444TB

Quant Time: Apr 01 03:29:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

