

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021695.D
 Acq On : 31 Mar 2021 18:58
 Operator : JC/MD
 Sample : PB135444ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135444ZHE#01

Quant Time: Apr 01 03:29:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	105238	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50014	51.90	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.80%	
35) Dibromofluoromethane	5.464	113	39140	52.36	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.72%	
50) Toluene-d8	8.694	98	138399	51.95	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.90%	
62) 4-Bromofluorobenzene	11.118	95	55827	53.90	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.80%	
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	18758	51.48	ug/l #	82
20) Methylene Chloride	2.835	84	5145	7.24	ug/l #	77
25) 2-Butanone	4.641	43	3353	6.07	ug/l #	87
43) Isopropyl Acetate	6.411	43	8625	4.46	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021695.D
Acq On : 31 Mar 2021 18:58
Operator : JC/MD
Sample : PB135444ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135444ZHE#01

Quant Time: Apr 01 03:29:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

