

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
 Data File : VX021697.D
 Acq On : 31 Mar 2021 19:45
 Operator : JC/MD
 Sample : M1872-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-GIS

Quant Time: Apr 01 03:30:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	58036	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117427	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50512	52.31	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.62%	
35) Dibromofluoromethane	5.470	113	39134	51.39	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.78%	
50) Toluene-d8	8.694	98	142280	52.43	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.86%	
62) 4-Bromofluorobenzene	11.118	95	57711	54.70	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.40%	
Target Compounds						
						Qvalue
16) Acetone	2.423	43	12384	33.92	ug/l #	85
20) Methylene Chloride	2.834	84	2824	3.96	ug/l #	81
25) 2-Butanone	4.640	43	3998	7.22	ug/l	92
43) Isopropyl Acetate	6.411	43	8907	4.52	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033121\
Data File : VX021697.D
Acq On : 31 Mar 2021 19:45
Operator : JC/MD
Sample : M1872-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-GIS

Quant Time: Apr 01 03:30:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

