

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027837.D
 Acq On : 31 Mar 2022 10:54
 Operator : JC/MD
 Sample : VX0331MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0331MBL01

Quant Time: Mar 31 15:17:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77496	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	134980	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53525	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63321	63.327	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	126.660%
35) Dibromofluoromethane	5.391	113	47479	54.494	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.980%
50) Toluene-d8	8.647	98	159377	51.991	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.980%
62) 4-Bromofluorobenzene	11.085	95	62694	48.125	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027837.D
Acq On : 31 Mar 2022 10:54
Operator : JC/MD
Sample : VX0331MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0331MBL01

Quant Time: Mar 31 15:17:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

