

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027838.D
 Acq On : 31 Mar 2022 11:21
 Operator : JC/MD
 Sample : VX0331WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0331WBL01

Quant Time: Apr 01 04:05:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75172	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	126305	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46084	47.513	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.020%
35) Dibromofluoromethane	5.391	113	42381	51.984	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.960%
50) Toluene-d8	8.653	98	137705	48.006	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.079	95	54101	44.382	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027838.D
Acq On : 31 Mar 2022 11:21
Operator : JC/MD
Sample : VX0331WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0331WBL01

Quant Time: Apr 01 04:05:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

