

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027841.D
 Acq On : 31 Mar 2022 12:33
 Operator : JC/MD
 Sample : N2041-17DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20220317DL

Quant Time: Apr 01 04:07:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	124388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52795	51.129	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.260%
35) Dibromofluoromethane	5.391	113	39901	49.696	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.400%
50) Toluene-d8	8.653	98	140519	49.742	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.480%
62) 4-Bromofluorobenzene	11.086	95	50217	41.830	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.660%
Target Compounds						
44) Trichloroethene	7.129	130	41281	49.567	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027841.D
Acq On : 31 Mar 2022 12:33
Operator : JC/MD
Sample : N2041-17DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20220317DL

Quant Time: Apr 01 04:07:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

