

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027843.D
 Acq On : 31 Mar 2022 13:20
 Operator : JC/MD
 Sample : N2041-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-20220318

Quant Time: Apr 01 04:25:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75694	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	129792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47065	48.190	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.380%
35) Dibromofluoromethane	5.391	113	44188	52.744	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.480%
50) Toluene-d8	8.653	98	149844	50.835	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.660%
62) 4-Bromofluorobenzene	11.085	95	57814	46.153	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.300%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3054	7.066	ug/l	# 78
25) 2-Butanone	4.581	43	868	1.380	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027843.D
Acq On : 31 Mar 2022 13:20
Operator : JC/MD
Sample : N2041-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-20220318

Quant Time: Apr 01 04:25:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

