

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027844.D
 Acq On : 31 Mar 2022 13:44
 Operator : JC/MD
 Sample : N2041-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-20220318

Quant Time: Apr 01 04:08:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	72758	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	129930	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	61237	65.231	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 130.460%	
35) Dibromofluoromethane	5.391	113	44465	53.019	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.040%	
50) Toluene-d8	8.653	98	159210	53.955	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.900%	
62) 4-Bromofluorobenzene	11.085	95	59995	47.844	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.680%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2959	7.122	ug/l	# 86
25) 2-Butanone	4.587	43	909	1.504	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027844.D
Acq On : 31 Mar 2022 13:44
Operator : JC/MD
Sample : N2041-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-20220318

Quant Time: Apr 01 04:08:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

