

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX033122\  
 Data File : VX027847.D  
 Acq On : 31 Mar 2022 14:55  
 Operator : JC/MD  
 Sample : N2157-01  
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MHD

Quant Time: Mar 31 15:30:16 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032922W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 30 04:39:27 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 70681    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 117640   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.061 | 117   | 107089   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 48830    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 47862    | 52.481   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 104.960% |          |
| 35) Dibromofluoromethane    | 5.397  | 113   | 37207    | 48.999   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 98.000%  |          |
| 50) Toluene-d8              | 8.653  | 98    | 138076   | 51.681   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 103.360% |          |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 56951    | 50.161   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 100.320% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX033122\  
Data File : VX027847.D  
Acq On : 31 Mar 2022 14:55  
Operator : JC/MD  
Sample : N2157-01  
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MHD

Quant Time: Mar 31 15:30:16 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X032922W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 30 04:39:27 2022  
Response via : Initial Calibration

