

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027850.D
 Acq On : 31 Mar 2022 16:06
 Operator : JC/MD
 Sample : N2041-21DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20220318DL

Quant Time: Apr 01 04:08:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	72818	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	134591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58747	62.527	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	125.060%
35) Dibromofluoromethane	5.391	113	44569	51.302	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.600%
50) Toluene-d8	8.653	98	158057	51.709	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.420%
62) 4-Bromofluorobenzene	11.079	95	62282	47.947	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.900%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	2663	3.906	ug/l	97
12) 1,1-Dichloroethene	2.319	96	344	0.550	ug/l	# 67
27) cis-1,2-Dichloroethene	4.495	96	683	0.834	ug/l	84
30) Chloroform	5.111	83	633	0.440	ug/l	88
44) Trichloroethene	7.129	130	60713	67.373	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027850.D
Acq On : 31 Mar 2022 16:06
Operator : JC/MD
Sample : N2041-21DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20220318DL

Quant Time: Apr 01 04:08:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

