

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027851.D
 Acq On : 31 Mar 2022 16:29
 Operator : JC/MD
 Sample : N2041-25DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20220318DL

Quant Time: Apr 01 04:09:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	73404	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	133464	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55802	58.918	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.840%
35) Dibromofluoromethane	5.391	113	43764	50.801	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.600%
50) Toluene-d8	8.653	98	152496	50.311	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.085	95	61184	47.500	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.000%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	47110	52.720	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027851.D
Acq On : 31 Mar 2022 16:29
Operator : JC/MD
Sample : N2041-25DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20220318DL

Quant Time: Apr 01 04:09:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

