

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027856.D
 Acq On : 31 Mar 2022 18:28
 Operator : JC/MD
 Sample : N2074-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20220319

Quant Time: Apr 01 04:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72185	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	129641	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52665	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61065	65.564	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 131.120%		
35) Dibromofluoromethane	5.391	113	44941	53.706	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 107.420%		
50) Toluene-d8	8.653	98	156760	53.243	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 106.480%		
62) 4-Bromofluorobenzene	11.085	95	62534	49.979	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 99.960%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	253671	375.299	ug/l	96
12) 1,1-Dichloroethene	2.325	96	1836	2.960	ug/l	86
16) Acetone	2.386	43	751	1.822	ug/l #	85
27) cis-1,2-Dichloroethene	4.489	96	629	0.775	ug/l	86
44) Trichloroethene	7.129	130	22116	25.479	ug/l	90
64) Tetrachloroethene	9.275	164	8634	10.489	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027856.D
Acq On : 31 Mar 2022 18:28
Operator : JC/MD
Sample : N2074-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20220319

Quant Time: Apr 01 04:10:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

