

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027859.D
 Acq On : 31 Mar 2022 19:38
 Operator : JC/MD
 Sample : N2074-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20220319

Quant Time: Apr 01 04:11:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	72943	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	132834	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120805	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51651	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62515	66.42	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	132.84%
35) Dibromofluoromethane	5.391	113	45444	53.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.00%
50) Toluene-d8	8.653	98	158653	52.59	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.18%
62) 4-Bromofluorobenzene	11.079	95	60745	47.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.76%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.167	85	212	0.32	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	2581	3.78	ug/l	96
12) 1,1-Dichloroethene	2.325	96	643	1.03	ug/l	# 67
19) Methyl tert-butyl Ether	3.117	73	2492	1.04	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	3537	4.31	ug/l	94
30) Chloroform	5.105	83	1926	1.34	ug/l	# 79
44) Trichloroethene	7.129	130	207182	232.95	ug/l	88
49) 1,4-Dioxane	7.665	88	649	29.24	ug/l	98
64) Tetrachloroethene	9.275	164	12521	15.29	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027859.D
Acq On : 31 Mar 2022 19:38
Operator : JC/MD
Sample : N2074-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20220319

Quant Time: Apr 01 04:11:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

