

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027866.D
 Acq On : 31 Mar 2022 23:16
 Operator : JC/MD
 Sample : VX0331WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0331WBL02

Quant Time: Apr 01 05:06:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74488	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	122708	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56351	58.632	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 117.260%	
35) Dibromofluoromethane	5.391	113	39287	49.602	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.200%	
50) Toluene-d8	8.653	98	143464	51.480	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.960%	
62) 4-Bromofluorobenzene	11.079	95	62307	52.612	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027866.D
Acq On : 31 Mar 2022 23:16
Operator : JC/MD
Sample : VX0331WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0331WBL02

Quant Time: Apr 01 05:06:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

