

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027869.D
 Acq On : 01 Apr 2022 00:26
 Operator : JC/MD
 Sample : N2074-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220320

Quant Time: Apr 01 05:08:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78828	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	138144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63879	62.805	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 125.620%	
35) Dibromofluoromethane	5.391	113	47169	52.899	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.800%	
50) Toluene-d8	8.653	98	164830	52.538	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	11.085	95	64281	48.213	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.420%	
Target Compounds						
16) Acetone	2.379	43	924	2.053	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027869.D
Acq On : 01 Apr 2022 00:26
Operator : JC/MD
Sample : N2074-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220320

Quant Time: Apr 01 05:08:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

