

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027870.D
 Acq On : 01 Apr 2022 00:49
 Operator : JC/MD
 Sample : N2074-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20220321

Quant Time: Apr 01 05:08:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76708	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	134506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63233	63.888	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 127.780%	
35) Dibromofluoromethane	5.397	113	45998	52.981	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.960%	
50) Toluene-d8	8.653	98	159242	52.130	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.260%	
62) 4-Bromofluorobenzene	11.079	95	63169	48.661	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.320%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	939	2.144	ug/l	89
44) Trichloroethene	7.129	130	1810	2.010	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027870.D
Acq On : 01 Apr 2022 00:49
Operator : JC/MD
Sample : N2074-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20220321

Quant Time: Apr 01 05:08:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

