

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
 Data File : VX027872.D
 Acq On : 01 Apr 2022 01:37
 Operator : JC/MD
 Sample : N2074-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW181S-20220321

Quant Time: Apr 01 05:09:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75891	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	131163	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62419	63.745	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 127.480%	
35) Dibromofluoromethane	5.391	113	45402	53.627	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.260%	
50) Toluene-d8	8.653	98	159194	53.442	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.880%	
62) 4-Bromofluorobenzene	11.085	95	63846	50.436	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.880%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	480	0.736	ug/l	85
16) Acetone	2.380	43	845	1.950	ug/l #	80
27) cis-1,2-Dichloroethene	4.495	96	875	1.025	ug/l	93
30) Chloroform	5.111	83	1073	0.715	ug/l #	79
44) Trichloroethene	7.135	130	1978	2.252	ug/l	86
64) Tetrachloroethene	9.275	164	1126	1.351	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033122\
Data File : VX027872.D
Acq On : 01 Apr 2022 01:37
Operator : JC/MD
Sample : N2074-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW181S-20220321

Quant Time: Apr 01 05:09:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

