

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034865.D
 Acq On : 31 Mar 2023 10:00
 Operator : JC/MD
 Sample : VX0331WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0331WBL01

Quant Time: Apr 01 05:21:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	237987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	412265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164384	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	170336	46.800	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.600%
35) Dibromofluoromethane	5.379	113	128814	47.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.480%
50) Toluene-d8	8.647	98	494527	49.134	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	189077	48.546	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034865.D
Acq On : 31 Mar 2023 10:00
Operator : JC/MD
Sample : VX0331WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0331WBL01

Quant Time: Apr 01 05:21:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

