

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034881.D
 Acq On : 31 Mar 2023 16:21
 Operator : JC/MD
 Sample : 02149-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 03 01:20:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	225529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	391644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	351817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154332	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165056	47.854	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.700%
35) Dibromofluoromethane	5.385	113	122930	47.456	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	8.646	98	508669	53.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.400%
62) 4-Bromofluorobenzene	11.079	95	166542	45.011	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034881.D
Acq On : 31 Mar 2023 16:21
Operator : JC/MD
Sample : 02149-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 03 01:20:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

