

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034898.D
 Acq On : 01 Apr 2023 03:27
 Operator : JC/MD
 Sample : 02140-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-02DUP

Quant Time: Apr 03 03:19:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	368139	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	317993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155265	47.318	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.640%		
35) Dibromofluoromethane	5.385	113	115172	47.300	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.600%		
50) Toluene-d8	8.647	98	440997	49.068	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.140%		
62) 4-Bromofluorobenzene	11.079	95	146227	42.044	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.080%		
Target Compounds						Qvalue
30) Chloroform	5.092	83	5275	1.109	ug/l	88
94) Hexachlorobutadiene	13.725	225	238	0.223	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034898.D
Acq On : 01 Apr 2023 03:27
Operator : JC/MD
Sample : 02140-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-02DUP

Quant Time: Apr 03 03:19:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

