

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
 Data File : VX034900.D
 Acq On : 01 Apr 2023 04:14
 Operator : JC/MD
 Sample : 02137-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-6

Quant Time: Apr 03 03:20:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	264912	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	465744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191251	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	213336	52.657	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.320%
35) Dibromofluoromethane	5.385	113	145499	47.232	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.460%
50) Toluene-d8	8.647	98	553684	48.695	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.079	95	215380	48.949	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX033123\
Data File : VX034900.D
Acq On : 01 Apr 2023 04:14
Operator : JC/MD
Sample : 02137-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-6

Quant Time: Apr 03 03:20:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

